



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **LKF – Laboratorium fuer Klinische Forschung**, using **Roche cobas c 501** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:

Instrument: <b>Cobas c 501</b>	Calibrator Lot: <b>229753</b>	Application: <b>(V-HbA1c/V-HB) * 91.5 + 2.15</b>
Reagent Lot: <b>317514, 327117</b>	Calibrator Assigned Values: <b>0.131, 0.335, 0.665, 0.927, 1.266, 1.809 mmol/L HbA1c; 1.422, 7.666 mmol/L Hb</b>	

Date of Certification: August 1, 2018

Certification Expires: August 1, 2019

*Daniel Sacks*

NGSP Steering Committee Chair

*Randie R Little PhD.*

NGSP Network Coordinator

*Chf*

SRL director/ supervisor